

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
-10 °C	-10.249 °C	-10.0 °C	0.249 °C	± 2.0 °C	± 4.0E-01 °C
2 °C	2.266 °C	2.5 °C	0.234 °C	± 2.0 °C	± 4.0E-01 °C
10 °C	10.007 °C	10.4 °C	0.393 °C	± 2.0 °C	± 4.0E-01 °C