

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-10 °C	-10.249 °C	-10.6 °C	-0.351 °C	± 0.5 °C	± 4.0E-01 °C
2 °C	2.266 °C	2.1 °C	-0.166 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.007 °C	9.9 °C	-0.107 °C	± 0.5 °C	± 4.0E-01 °C