

| <b>Nominal (Nominal)</b> | <b>Patrón (Standard)</b> | <b>UBP (UUT)</b> | <b>Error (Error)</b> | <b>Exactitud (Accuracy)</b> | <b>Incertidumbre (Uncertainty)</b> |
|--------------------------|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|
| -10 °C                   | -10.249 °C               | -10.2 °C         | 0.049 °C             | ± 2.0 °C                    | ± 4.0E-01 °C                       |
| 2 °C                     | 2.266 °C                 | 2.2 °C           | -0.066 °C            | ± 2.0 °C                    | ± 4.0E-01 °C                       |
| 10 °C                    | 10.007 °C                | 10.0 °C          | -0.007 °C            | ± 2.0 °C                    | ± 4.0E-01 °C                       |