

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.7	0.69920	1.070	± 1.7	± 4.3E-01 mm					
0 a 0.8	0.79910	1.170	± 1.8	± 4.3E-01 mm					
0 a 0.9	0.89900	1.280	± 1.9	± 4.3E-01 mm					
0 a 1	0.99890	1.380	± 2	± 4.8E-01 mm					
0 a 1.1	1.09880	1.480	± 2	± 4.8E-01 mm					
0 a 1.2	1.19870	1.590	± 2.1	± 4.8E-01 mm					
0 a 1.3	1.29860	1.690	± 2.2	± 4.8E-01 mm					
0 a 1.4	1.39850	1.800	± 2.3	± 4.8E-01 mm					
0 a 1.5	1.49840	1.900	± 2.4	± 4.8E-01 mm					
0 a 1.6	1.59830	2.010	± 2.4	± 4.8E-01 mm					
0 a 1.7	1.69820	2.110	± 2.5	± 4.8E-01 mm					
0 a 1.8	1.79810	2.220	± 2.6	± 4.8E-01 mm					
0 a 1.9	1.89800	2.330	± 2.7	± 4.8E-01 mm					
0 a 2	1.99790	2.430	± 2.8	± 4.8E-01 mm					