

Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
Profundidad / (Depth)	1 mm	1.00 mm	0 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	2 mm	2.01 mm	10 µm	± 30 µm	± 5.8E+00 µm
Profundidad / (Depth)	3 mm	3.00 mm	0 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	5 mm	5.00 mm	0 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	10 mm	10.01 mm	10 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	12 mm	12.00 mm	0 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	15 mm	15.00 mm	0 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	20 mm	20.01 mm	10 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	25 mm	25.00 mm	0 µm	± 30 µm	± 6.1E+00 µm