

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
99.9135 µl	100 µl	0.087 µl	± 0.8 µl	±	5.3E-02 µl
499.9182 µl	500 µl	0.082 µl	± 4 µl	±	5.9E-02 µl
999.736 µl	1000 µl	0.264 µl	± 8 µl	±	7.8E-02 µl

