

| Nominal (<i>Nominal</i>) | Patrón (<i>Standard</i>) | UBP (<i>UUT</i>) | Error (<i>Error</i>) | Exactitud (<i>Accuracy</i>) | Incertidumbre (<i>Uncertainty</i>) |
|-----------------------------------|-----------------------------------|---------------------------|-------------------------------|--------------------------------------|---|
| 30 °C | 30.664 °C | 30.9 °C | 0.236 °C | ± 1.0 °C | ± 4.0E-01 °C |
| 35 %HR | 35.171 %HR | 33.0 %HR | -2.171 %HR | ± 5.0 %HR | ± 1.9E+00 %HR |
| 50 %HR | 50.130 %HR | 49.0 %HR | -1.130 %HR | ± 5.0 %HR | ± 1.9E+00 %HR |
| 70 %HR | 70.448 %HR | 69.0 %HR | -1.448 %HR | ± 5.0 %HR | ± 1.9E+00 %HR |