

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
30 °C	30.664 °C	30.5 °C	-0.164 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	35.171 %HR	35.0 %HR	-0.171 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.130 %HR	50.3 %HR	0.170 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	70.3 %HR	-0.148 %HR	± 3.0 %HR	± 1.9E+00 %HR