

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.664 °C	30.5 °C	-0.164 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	35.171 %HR	34.3 %HR	-0.871 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.130 %HR	51.3 %HR	1.170 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	70.9 %HR	0.452 %HR	± 3.0 %HR	± 1.9E+00 %HR