

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.664 °C	31.0 °C	0.336 °C	± 1.0 °C	± 4.0E-01 °C
35 %HR	35.171 %HR	37.0 %HR	1.829 %HR	± 4.0 %HR	± 1.9E+00 %HR
50 %HR	50.130 %HR	50.0 %HR	-0.130 %HR	± 4.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	71.0 %HR	0.552 %HR	± 4.0 %HR	± 1.9E+00 %HR