

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.664 °C	31.0 °C	0.336 °C	± 1.0 °C	± 4.0E-01 °C
35 %HR	35.171 %HR	37.0 %HR	1.829 %HR	± 4.0 %HR	± 1.9E+00 %HR
50 %HR	50.130 %HR	53.0 %HR	2.870 %HR	± 4.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	72.0 %HR	1.552 %HR	± 4.0 %HR	± 1.9E+00 %HR