

---

<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
43.964 ml	44.000 ml	0.036 ml	± 0.05 ml	± 1.4E+00 µl

---