

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.222 °C	15.2 °C	-0.022 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.550 °C	25.5 °C	-0.050 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.670 °C	32.7 °C	0.030 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	20.345 %HR	22.3 %HR	1.955 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.263 %HR	50.9 %HR	0.637 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.141 %HR	68.0 %HR	-2.141 %HR	± 5.0 %HR	± 1.9E+00 %HR