

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|--------------------------|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|
| 15 °C | 15.222 °C | 15.5 °C | 0.278 °C | ± 0.5 °C | ± 4.0E-01 °C |
| 25 °C | 25.550 °C | 25.7 °C | 0.150 °C | ± 0.5 °C | ± 4.0E-01 °C |
| 32 °C | 32.670 °C | 32.5 °C | -0.170 °C | ± 0.5 °C | ± 4.0E-01 °C |
| 20 %HR | 20.345 %HR | 22.3 %HR | 1.955 %HR | ± 5.0 %HR | ± 1.9E+00 %HR |
| 50 %HR | 50.263 %HR | 49.9 %HR | -0.363 %HR | ± 5.0 %HR | ± 1.9E+00 %HR |
| 70 %HR | 70.141 %HR | 68.8 %HR | -1.341 %HR | ± 5.0 %HR | ± 1.9E+00 %HR |