

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.222 °C	15.3 °C	0.078 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.550 °C	25.8 °C	0.250 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.670 °C	32.9 °C	0.230 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	20.345 %HR	21.5 %HR	1.155 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.263 %HR	51.2 %HR	0.937 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.141 %HR	70.2 %HR	0.059 %HR	± 5.0 %HR	± 1.9E+00 %HR