

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.222 °C	15.1 °C	-0.122 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.550 °C	25.7 °C	0.150 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.670 °C	32.5 °C	-0.170 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	20.345 %HR	21.6 %HR	1.255 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.263 %HR	48.2 %HR	-2.063 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.141 %HR	71.2 %HR	1.059 %HR	± 5.0 %HR	± 1.9E+00 %HR