

<b>Nominal</b> ( <i>Nominal</i> )	<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UUT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
15 °C	15.222 °C	15.4 °C	0.178 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.550 °C	25.6 °C	0.050 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.670 °C	32.6 °C	-0.070 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	20.345 %HR	22.7 %HR	2.355 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.263 %HR	52.2 %HR	1.937 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.141 %HR	72.0 %HR	1.859 %HR	± 5.0 %HR	± 1.9E+00 %HR