

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.222 °C	15.1 °C	-0.122 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.550 °C	25.5 °C	-0.050 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.670 °C	32.5 °C	-0.170 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	20.345 %HR	21.9 %HR	1.555 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.263 %HR	53.5 %HR	3.237 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.141 %HR	70.8 %HR	0.659 %HR	± 5.0 %HR	± 1.9E+00 %HR