

<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
18.9277 l	18.9271 l	-0.0006 l	± 0.0038 l	± 7.3E-01 µl
5.0002 gal	5.0000 gal	-0.0002 gal	± 0.0010 gal	± 2.0E-07 gal

