

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.731 °C	29.5 °C	-0.231 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.417 %HR	33.4 %HR	2.983 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.527 %HR	51.6 %HR	2.073 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	76.051 %HR	73.1 %HR	-2.951 %HR	± 3.0 %HR	± 1.9E+00 %HR