

<b>Nominal</b> ( <i>Nominal</i> )	<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UUT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
30 °C	29.731 °C	29.5 °C	-0.231 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.417 %HR	33.0 %HR	2.583 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.527 %HR	51.1 %HR	1.573 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	76.051 %HR	74.2 %HR	-1.851 %HR	± 3.0 %HR	± 1.9E+00 %HR