

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.731 °C	29.6 °C	-0.131 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.417 %HR	33.4 %HR	2.983 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.527 %HR	50.3 %HR	0.773 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	76.051 %HR	73.9 %HR	-2.151 %HR	± 3.0 %HR	± 1.9E+00 %HR