

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.453 °C	30.3 °C	-0.153 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.558 %HR	31.3 %HR	1.742 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.006 %HR	47.3 %HR	-1.706 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.100 %HR	72.8 %HR	-2.300 %HR	± 3.0 %HR	± 1.9E+00 %HR