

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
30 °C	31.185 °C	31.2 °C	0.015 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.560 %HR	33 %HR	2.440 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.190 %HR	51 %HR	1.810 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	74.313 %HR	76 %HR	1.687 %HR	± 5.0 %HR	± 1.9E+00 %HR