

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	31.185 °C	31.7 °C	0.515 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.560 %HR	34 %HR	3.440 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.190 %HR	53 %HR	3.810 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	74.313 %HR	79 %HR	4.687 %HR	± 5.0 %HR	± 1.9E+00 %HR