

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	31.185 °C	31.7 °C	0.515 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.560 %HR	31.0 %HR	0.440 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.190 %HR	52.0 %HR	2.810 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	74.313 %HR	79.0 %HR	4.687 %HR	± 5.0 %HR	± 1.9E+00 %HR