

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	31.185 °C	30.7 °C	-0.485 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.560 %HR	29.0 %HR	-1.560 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.190 %HR	49.0 %HR	-0.190 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	74.313 %HR	76.0 %HR	1.687 %HR	± 5.0 %HR	± 1.9E+00 %HR