

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
30 °C	31.185 °C	31.7 °C	0.515 °C	± 0.6 °C	± 4.0E-01 °C
30 %HR	30.560 %HR	33.4 %HR	2.840 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.190 %HR	51.1 %HR	1.910 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	74.313 %HR	76.1 %HR	1.787 %HR	± 3.0 %HR	± 1.9E+00 %HR