
Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
24 °C	24.146 °C	23.9 °C	-0.296 °C	± 1.0 °C	± 4.0E-01 °C
53 %HR	52.932 %HR	55.0 %HR	2.068 %HR	± 3.0 %HR	± 1.9E+00 %HR