

---

<b>Nominal</b> ( <i>Nominal</i> )	<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UUT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
24 °C	24.146 °C	23.9 °C	-0.296 °C	± 1.0 °C	± 4.0E-01 °C
53 %HR	52.932 %HR	55.0 %HR	2.068 %HR	± 3.0 %HR	± 1.9E+00 %HR