

**Clase de Exactitud (Accuracy Class):** Clase II (Class II)

---

<b>Valor UBP</b>	<b>Valor Patrón</b>	<b>Error</b>	<b>EMP</b>	<b>Incertidumbre</b>	<b>Valor UBP</b>	<b>Valor Patrón</b>	<b>Error</b>	<b>EMP</b>	<b>Incertidumbre</b>
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.1	0.09989	0.110	± 0.6	± 4.3E-01 mm					
0 a 0.2	0.19986	0.140	± 0.6	± 4.3E-01 mm					
0 a 0.3	0.29984	0.160	± 0.7	± 4.3E-01 mm					
0 a 0.4	0.39973	0.270	± 0.7	± 4.3E-01 mm					
0 a 0.5	0.49972	0.280	± 0.8	± 4.3E-01 mm					
0 a 0.6	0.59966	0.340	± 0.8	± 4.3E-01 mm					
0 a 0.7	0.69956	0.440	± 0.8	± 4.3E-01 mm					
0 a 0.8	0.79959	0.410	± 0.9	± 4.3E-01 mm					
0 a 0.9	0.89955	0.450	± 0.9	± 4.3E-01 mm					
0 a 1	0.99939	0.610	± 1	± 4.3E-01 mm					

---