

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	14.901 °C	15.2 °C	0.299 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.008 °C	25.3 °C	0.292 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.183 °C	30.4 °C	0.217 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.560 %HR	21.0 %HR	-9.560 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.190 %HR	40.4 %HR	-8.790 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.223 %HR	62.3 %HR	-7.923 %HR	± 5.0 %HR	± 1.9E+00 %HR