

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	14.901 °C	15.4 °C	0.499 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.008 °C	25.5 °C	0.492 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.183 °C	30.8 °C	0.617 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.560 %HR	30.6 %HR	0.040 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.190 %HR	48.5 %HR	-0.690 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.223 %HR	69.8 %HR	-0.423 %HR	± 5.0 %HR	± 1.9E+00 %HR