

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	14.901 °C	15.1 °C	0.199 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.008 °C	25.2 °C	0.192 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.183 °C	30.3 °C	0.117 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.560 %HR	32.5 %HR	1.940 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.190 %HR	51.9 %HR	2.710 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.223 %HR	68.8 %HR	-1.423 %HR	± 3.0 %HR	± 1.9E+00 %HR