

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.842 °C	29.6 °C	-0.242 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	29.762 %HR	32 %HR	2.238 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.877 %HR	55 %HR	4.123 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	74.302 %HR	79 %HR	4.698 %HR	± 5.0 %HR	± 1.9E+00 %HR