

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
30 °C	29.842 °C	30 °C	0.158 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	29.762 %HR	- %HR	- %HR	± 5.0 %HR	± - -
50 %HR	48.842 %HR	43 %HR	-5.842 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.366 %HR	62 %HR	-8.366 %HR	± 5.0 %HR	± 1.9E+00 %HR