

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.842 °C	30 °C	0.158 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	29.762 %HR	- %HR	- %HR	± 5.0 %HR	± - -
50 %HR	50.079 %HR	38 %HR	-12.079 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.366 %HR	62 %HR	-8.366 %HR	± 5.0 %HR	± 1.9E+00 %HR