

Patrón (Standard)	UBP (UT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)			
3601.454 s	3602 s	0.546 s	0.0152%	±	0.01%	±	5.6E-02 s
3601.489 s	3602 s	0.511 s	0.0142%	±	0.01%	±	5.6E-02 s
3602.498 s	3603 s	0.502 s	0.0139%	±	0.01%	±	5.6E-02 s