

Patrón (Standard)	UBP (UT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
3603.843 s	3604 s	0.157 s	0.0044%	± 0.01%	± 5.6E-02 s
3602.785 s	3603 s	0.215 s	0.0060%	± 0.01%	± 5.6E-02 s
3603.948 s	3604 s	0.052 s	0.0014%	± 0.01%	± 5.6E-02 s