

<b>Nominal</b> ( <i>Nominal</i> )	<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UUT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
15 °C	14.690 °C	14.5 °C	-0.190 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.882 °C	29.7 °C	-0.182 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	29.670 %HR	31.6 %HR	1.930 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.231 %HR	52.2 %HR	2.969 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	69.568 %HR	71.8 %HR	2.232 %HR	± 3.0 %HR	± 1.9E+00 %HR