

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.842 °C	29.8 °C	-0.042 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	29.762 %HR	32.2 %HR	2.438 %HR	± 4.0 %HR	± 1.9E+00 %HR
50 %HR	50.079 %HR	52.0 %HR	1.921 %HR	± 4.0 %HR	± 1.9E+00 %HR
75 %HR	69.343 %HR	73.1 %HR	3.757 %HR	± 4.0 %HR	± 1.9E+00 %HR