

<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>	
99.8132 µl	100 µl	0.187 µl	± 0.8 µl	±	5.3E-02 µl
500.0074 µl	500 µl	-0.007 µl	± 4 µl	±	5.9E-02 µl
1000.052 µl	1000 µl	-0.052 µl	± 8 µl	±	7.8E-02 µl