

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
0.0625 in	0.063 in	0.001 in	± 0.003 in	±	5.8E-04 in
0.100 in	0.100 in	0.000 in	± 0.003 in	±	5.8E-04 in
0.125 in	0.125 in	0.000 in	± 0.003 in	±	5.8E-04 in
0.200 in	0.200 in	0.000 in	± 0.003 in	±	5.9E-04 in
0.250 in	0.250 in	0.000 in	± 0.003 in	±	6.0E-04 in
0.300 in	0.299 in	-0.001 in	± 0.003 in	±	6.1E-04 in
0.500 in	0.498 in	-0.002 in	± 0.003 in	±	6.7E-04 in
1.000 in	0.997 in	-0.003 in	± 0.003 in	±	9.1E-04 in