

<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
0.0625 in	0.063 in	0.001 in		± 5.8E-04 in
0.100 in	0.100 in	0.000 in		± 5.8E-04 in
0.125 in	0.125 in	0.000 in		± 5.8E-04 in
0.200 in	0.200 in	0.000 in		± 5.9E-04 in
0.250 in	0.250 in	0.000 in		± 6.0E-04 in
0.300 in	0.299 in	-0.001 in		± 6.1E-04 in
0.500 in	0.498 in	-0.002 in		± 6.7E-04 in
1.000 in	0.997 in	-0.003 in		± 9.1E-04 in