

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
1.005 mm	1.008 mm	3 μm	$\pm 50.25 \mu\text{m}$	\pm	5.8E-01 μm
1.5 mm	1.502 mm	2 μm	$\pm 75. \mu\text{m}$	\pm	5.8E-01 μm
2 mm	2.002 mm	2 μm	$\pm 100 \mu\text{m}$	\pm	5.8E-01 μm
4 mm	4.008 mm	8 μm	$\pm 200 \mu\text{m}$	\pm	5.9E-01 μm
5 mm	5.010 mm	10 μm	$\pm 250 \mu\text{m}$	\pm	5.8E-01 μm