

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
25 °C	24.848 °C	24.9 °C	0.052 °C	± 0.5 °C	± 4.0E-01 °C
# 25 °C	24.848 °C	24.8 °C	-0.048 °C	± 0.5 °C	± 7.4E-02 °C
# 50 °C	49.917 °C	50.1 °C	0.183 °C	± 0.5 °C	± 7.4E-02 °C
# 70 °C	69.972 °C	70.2 °C	0.228 °C	± 0.5 °C	± 7.4E-02 °C
25 %HR	23.413 %HR	25.0 %HR	1.587 %HR	± 3 %HR	± 1.9E+00 %HR
50 %HR	48.975 %HR	51.1 %HR	2.125 %HR	± 3 %HR	± 1.9E+00 %HR
70 %HR	69.425 %HR	72.2 %HR	2.775 %HR	± 3 %HR	± 1.9E+00 %HR