

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
# 5 RPM	5 RPM	0 RPM	± 0.103 RPM	±	5.8E-02 RPM
# 30 RPM	30 RPM	0 RPM	± 0.115 RPM	±	5.8E-02 RPM
60 RPM	59.9 RPM	-0.1 RPM	± 0.13 RPM	±	5.8E-02 RPM
100 RPM	99 RPM	-1 RPM	± 0.15 RPM	±	8.0E-01 RPM
1500 RPM	1499 RPM	-1 RPM	± 2 RPM	±	8.4E-01 RPM
10000 RPM	9999 RPM	-1 RPM	± 6 RPM	±	8.0E-01 RPM
60000 RPM	59997 RPM	-3 RPM	± 31 RPM	±	8.4E-01 RPM