

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|--------------------------|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|
| 30 °C | 31.827 °C | 31.2 °C | -0.627 °C | ± 0.5 °C | ± 4.0E-01 °C |
| 30 %HR | 31.541 %HR | 33.7 %HR | 2.159 %HR | ± 3.0 %HR | ± 1.9E+00 %HR |
| 50 %HR | 49.057 %HR | 51.2 %HR | 2.143 %HR | ± 3.0 %HR | ± 1.9E+00 %HR |
| 75 %HR | 75.396 %HR | 74.3 %HR | -1.096 %HR | ± 3.0 %HR | ± 1.9E+00 %HR |