

<b>Patrón (Standard)</b>	<b>UBP (UT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>	
12.1 RPM	12 RPM	-0.1 RPM	± 1 RPM	±	9.4E-02 RPM
29.94 RPM	30 RPM	0.06 RPM	± 1 RPM	±	1.1E-01 RPM
60.04 RPM	60 RPM	-0.04 RPM	± 1 RPM	±	1.1E-01 RPM