

<b>Nominal</b> ( <i>Nominal</i> )	<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UUT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
2 °C	2.015 °C	2.4 °C	0.385 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.983 °C	5.3 °C	0.317 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.142 °C	8.6 °C	0.458 °C	± 0.5 °C	± 4.0E-01 °C